



# Certificate of Calibration

1000209034

Page 1 of 9



## Customer Information

Accurate Calibration & Repair  
1924 Pinnacle Drive  
Aurora, IL 60504

PO #: 2019006  
Reference #: 1946308Rg17025  
Account #: 00317  
SO #: 46308

## Instrument Identification

Instrument Id: **02.20.002725**

Location:

Noun: Handheld Calibrator

Model: Memocal 2000

Mfr: ERO Electronic

Serial #: 02.20.002725

Accuracy: See manufacturer's specifications

## Certification Information

Reason For Service: Calibration with Data

Technician: Appealle Bullock

Type Of Calibration: Accredited 17025

Cal Date: 27 MAR 19

As Found Condition: Out of Tolerance

Cal Due: 27 JUN 19

As Left Condition: In Tolerance, Adjusted

Temperature: 22.2 °C

Procedure: MFR Manual :

Humidity: 33.0 %

Technician Remarks: Datasheet reflects customer specified calibration points. Unit calibrated per AMS 2750 E.

- No sampling plan or other procedure was used for this calibration. Measurements and information on this certificate are valid at time of calibration only and any number of factors may cause calibration to drift out of tolerance prior to calibration due date.
- The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI).
- The expanded uncertainty of the measurement process has not exceeded 25% of the tolerance allowed for the individual characteristics measured, unless otherwise stated. The uncertainties are based on a 95% confidence level, K=2.
- (!) Designates that the expanded uncertainty of measurement does not meet the 95% confidence level.
- J.H. Metrology Co., Inc's Calibration Control System complies with applicable requirements of ANSI Z540-1-1994, ISO 9001, and ISO/IEC 17025-2005.
- The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without the written approval of J.H. Metrology Co., Inc.

Approved By:

President

Date: Mar 28, 2019

## Calibration Data

✓ In Tolerance    ✗ Out of Tolerance

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
mV Output						
-4 mV to 20 mV	-4.000	-4.00261	✓	-3.99962	✓	0.00060 mV
	0.000	-0.00072	✓	0.00093	✓	0.00059 mV

# Certificate of Calibration

1000209034

Page 2 of 9

✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>mV Output</b>						
	5.000	5.00057	✓	5.00055	✓	0.00060 mV
	10.000	10.00223	✓	10.00041	✓	0.00061 mV
	15.000	15.00385		15.00084	✓	0.00062 mV
	20.000	20.00623	✗	20.00113	✓	0.00063 mV
-40 mV to 200 mV	-40.00	-40.00537	✓	-39.99903	✓	0.0058 mV
	0.00	0.00554	✓	0.00566	✓	0.0058 mV
	50.00	50.01349	✓	50.00114	✓	0.0058 mV
	100.00	100.02383	✓	100.00178	✓	0.0059 mV
	150.00	150.03257	✓	150.00262	✓	0.0059 mV
	200.00	200.0458	✗	200.0076	✓	0.0060 mV
-400 mV to 2000 mV	-400.0	-400.0891	✓	-400.0318	✓	0.058 mV
	0.0	0.00552	✓	0.00609	✓	0.058 mV
	500.0	500.1041	✓	499.9777	✓	0.058 mV
	1000.0	1000.1692	✓	999.9643	✓	0.058 mV
	1500.0	1500.2761	✓	1499.9221	✓	0.059 mV
	2000.0	2000.438	✗	2000.058	✓	0.060 mV
-4000 mV to 20000 mV	-4000	-4001.149	✓	-4000.036	✓	0.58 mV
	0	-0.01354	✓	0.32639	✓	0.58 mV
	5000	5001.371	✓	5000.211	✓	0.58 mV
	10000	10002.142	✓	9999.798	✓	0.58 mV
	15000	15003.668	✓	15000.067	✓	0.59 mV
	20000	20005.05	✗	20000.42	✓	0.61 mV
<b>mV Measurement</b>						
-20 mV to 20 mV	-20.000	-19.995	✓	-20.000	✓	0.00090 mV
	0.000	-0.001	✓	-0.001	✓	0.00077 mV
	5.000	4.998	✓	5.000	✓	0.00080 mV
	10.000	9.995	✓	9.999	✓	0.00083 mV
	15.000	14.995	✓	15.000	✓	0.00086 mV
	20.000	19.994		20.001	✓	0.00090 mV
-200 mV to 200 mV	-200.00	-199.94		-199.99	✓	0.0063 mV
	0.00	0.00	✓	0.00	✓	0.0058 mV
	50.00	49.99	✓	50.00	✓	0.0059 mV
	100.00	99.98	✓	100.00	✓	0.0060 mV
	150.00	149.96	✓	150.00	✓	0.0061 mV

# Certificate of Calibration

1000209034

Page 3 of 9

✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found		As Left		Min	Max	Uncertainty
<b>mV Measurement</b>								
	200.00	199.95	✓	200.01	✓	199.94	200.06	0.0063 mV
-2000 mV to 2000 mV	-2000.0	-1999.3	✗	-1999.9	✓	-2000.6	-1999.4	0.060 mV
	0.0	0.0	✓	0.0	✓	-0.6	0.6	0.058 mV
	500.0	499.8	✓	500.0	✓	499.4	500.6	0.058 mV
	1000.0	999.7	✓	1000.0	✓	999.4	1000.6	0.059 mV
	1500.0	1499.6	✓	1500.1	✓	1499.4	1500.6	0.059 mV
	2000.0	1999.5	✓	2000.1	✓	1999.4	2000.6	0.060 mV
-20000 mV to 20000 mV	-20000	-19990	✗	-19999	✓	-20008	-19992	0.59 mV
	0	0	✓	0	✓	-8	8	0.58 mV
	5000	4997	✓	5000	✓	4992	5008	0.58 mV
	10000	9995	✓	10000	✓	9992	10008	0.58 mV
	15000	14993	✓	15001	✓	14992	15008	0.59 mV
	20000	19991	✗	20001	✓	19992	20008	0.59 mV
<b>mA Output</b>								
0 mA to 21 mA	0.000	0.00053152	✓	0.00019427	✓	-0.003	0.003	0.00058 mA
	1.000	1.0006224	✓	0.9999794	✓	0.997	1.003	0.00058 mA
	5.000	5.001945	✓	5.000742	✓	4.997	5.003	0.00060 mA
	10.000	10.003163	✗	10.000395	✓	9.997	10.003	0.00062 mA
	15.000	15.003976	✗	15.0000237	✓	14.997	15.003	0.00066 mA
	21.000	21.00593	✗	21.00064	✓	20.997	21.003	0.00072 mA
<b>mA Measurement</b>								
-20 mA to 20 mA	-20.000	-19.993	✗	-20.000	✓	-20.006	-19.994	0.0011 mA
	0.000	0.000	✓	0.000	✓	-0.006	0.006	0.00058 mA
	5.000	4.998	✓	5.000	✓	4.994	5.006	0.00063 mA
	10.000	9.997	✓	9.999	✓	9.994	10.006	0.00074 mA
	15.000	14.996	✓	15.000	✓	14.994	15.006	0.00087 mA
	20.000	19.995		20.000	✓	19.994	20.006	0.0011 mA
-130 mA to 130 mA	-130.00	-129.97	✓	-130.00	✓	-130.05	-129.95	0.0094 mA
	0.00	0.00	✓	0.00	✓	-0.05	0.05	0.0058 mA
	10.00	10.00	✓	10.00	✓	9.95	10.05	0.0058 mA
	50.00	49.99	✓	50.00	✓	49.95	50.05	0.0067 mA
	100.00	99.98	✓	100.01	✓	99.95	100.05	0.0082 mA
	130.00	129.98	✓	130.01	✓	129.95	130.05	0.0094 mA



# Certificate of Calibration

1000209034

Page 4 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found		As Left		Min	Max	Uncertainty
<b>RTD Pt. 100 Ohm Measurement (-200° to 850°C)</b>								
20.68 Ohms	-195.0	-195.0	✓	-195.0	✓	-195.6	-194.4	0.058 °C
60.26 Ohms	-100.0	-99.9	✓	-100.1	✓	-100.6	-99.4	0.058 °C
100.00 Ohms	0.0	0.2	✓	-0.1	✓	-0.6	0.6	0.058 °C
138.51 Ohms	100.0	100.3	✓	100.0	✓	99.4	100.6	0.058 °C
280.98 Ohms	500.0	500.8	✗	499.9	✓	499.4	500.6	0.059 °C
389.02 Ohms	845.0	846.2	✗	844.7	✓	844.1	845.9	0.060 °C
<b>Ohms Measurement</b>								
0 to 800 Ohms	0.0	0.0	✓	0.0	✓	-0.2	0.2	0.058 Ohms
	10.0	10.0	✓	10.0	✓	9.8	10.2	0.058 Ohms
	100.0	100.1	✓	100.0	✓	99.8	100.2	0.058 Ohms
	300.0	300.3	✗	300.0	✓	299.8	300.2	0.058 Ohms
	500.0	500.5	✗	499.9	✓	499.8	500.2	0.058 Ohms
	790.0	790.8	✗	789.9	✓	789.8	790.2	0.059 Ohms
<b>Thermocouple Measurement</b>								
Type E -328° to 1832°F	-328.0	-327.8	✓	-327.8	✓	-329.0	-327.0	0.061 °F
	-200.0	-200.0	✓	-200.0	✓	-201.0	-199.0	0.058 °F
	0.0	0.0	✓	0.0	✓	-1.0	1.0	0.058 °F
	32.0	32.0	✓	32.0	✓	31.0	33.0	0.058 °F
	200.0	199.8	✓	199.9	✓	199.0	201.0	0.058 °F
	400.0	399.7	✓	399.9	✓	399.0	401.0	0.058 °F
	600.0	599.9	✓	600.1	✓	599.0	601.0	0.058 °F
	800.0	799.9	✓	800.1	✓	799.0	801.0	0.058 °F
	1000.0	999.9	✓	1000.0	✓	999.0	1001.0	0.058 °F
	1200.0	1199.7	✓	1200.0	✓	1198.8	1201.2	0.058 °F
	1400.0	1399.6	✓	1400.0	✓	1398.6	1401.4	0.059 °F
	1600.0	1599.6	✓	1600.0	✓	1598.4	1601.6	0.059 °F
	1832.0	1831.6	✓	1832.2	✓	1830.2	1833.8	0.059 °F
Type J -328° to 2192°F	-328.0	-328.0	✓	-328.0	✓	-329.0	-327.0	0.058 °F
	-200.0	-200.0	✓	-200.0	✓	-201.0	-199.0	0.058 °F
	0.0	0.0	✓	0.0	✓	-1.0	1.0	0.058 °F
	32.0	32.0	✓	32.0	✓	31.0	33.0	0.058 °F
	200.0	199.8	✓	199.8	✓	199.0	201.0	0.058 °F
	400.0	399.7	✓	399.7	✓	399.0	401.0	0.058 °F
	600.0	599.9	✓	600.1	✓	599.0	601.0	0.058 °F



# Certificate of Calibration

1000209034

Page 5 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>Thermocouple Measurement</b>						
	800.0	799.9 ✓	800.1 ✓	799.0	801.0	0.058 °F
	1000.0	999.9 ✓	1000.0 ✓	999.0	1001.0	0.059 °F
	1200.0	1199.7 ✓	1200.0 ✓	1198.8	1201.2	0.059 °F
	1600.0	1599.6 ✓	1600.0 ✓	1598.4	1601.6	0.059 °F
	1800.0	1799.6 ✓	1800.0 ✓	1798.2	1801.8	0.059 °F
	2192.0	2191.3 ✓	2192.0 ✓	2189.8	2194.2	0.060 °F
Type K -328° to 2498°F	-328.0	-327.8 ✓	-328.0 ✓	-329.0	-327.0	0.079 °F
	-200.0	-200.0 ✓	-200.0 ✓	-201.0	-199.0	0.058 °F
	0.0	0.0 ✓	0.0 ✓	-1.0	1.0	0.058 °F
	32.0	32.0 ✓	32.0 ✓	31.0	33.0	0.058 °F
	200.0	199.9 ✓	199.9 ✓	199.0	201.0	0.058 °F
	400.0	399.7 ✓	399.7 ✓	399.0	401.0	0.058 °F
	600.0	599.7 ✓	599.9 ✓	599.0	601.0	0.058 °F
	800.0	799.7 ✓	799.9 ✓	799.0	801.0	0.058 °F
	1000.0	999.9 ✓	999.9 ✓	999.0	1001.0	0.059 °F
	1200.0	1199.8 ✓	1200.0 ✓	1198.8	1201.2	0.059 °F
	1400.0	1399.6 ✓	1400.0 ✓	1398.6	1401.4	0.059 °F
	1600.0	1599.6 ✓	1600.0 ✓	1598.4	1601.6	0.059 °F
	1800.0	1799.6 ✓	1800.0 ✓	1798.2	1801.8	0.059 °F
	2000.0	1999.2 ✓	1999.9 ✓	1998.0	2002.0	0.059 °F
	2200.0	2199.2 ✓	2199.9 ✓	2197.8	2202.2	0.062 °F
	2498.0	2497.1 ✓	2498.0 ✓	2495.5	2500.5	0.062 °F
Type T -328° to 752°F	-328.0	-327.8 ✓	-328.0 ✓	-329.0	-327.0	0.066 °F
	-200.0	-199.8 ✓	-200.0 ✓	-201.0	-199.0	0.059 °F
	0.0	0.0 ✓	0.0 ✓	-1.0	1.0	0.058 °F
	32.0	32.0 ✓	32.0 ✓	31.0	33.0	0.058 °F
	200.0	199.9 ✓	199.9 ✓	199.0	201.0	0.058 °F
	400.0	399.7 ✓	399.9 ✓	399.0	401.0	0.058 °F
	600.0	599.9 ✓	600.1 ✓	599.0	601.0	0.058 °F
	752.0	751.8 ✓	752.0 ✓	751.0	753.0	0.058 °F
Type R -58° to 3214°F	-50.0	-51.2 ✓	-50.6 ✓	-52.6	-47.4	0.065 °F
	0.0	-0.8 ✓	-0.8 ✓	-2.6	2.6	0.065 °F
	32.0	31.3 ✓	31.6 ✓	29.4	34.6	0.065 °F
	200.0	199.8 ✓	199.9 ✓	198.4	201.6	0.061 °F
	400.0	399.4 ✓	399.7 ✓	398.8	401.2	0.060 °F



# Certificate of Calibration

1000209034

Page 6 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>Thermocouple Measurement</b>						
	600.0	599.5 ✓	599.9 ✓	599.0	601.0	0.060 °F
	800.0	799.7 ✓	799.9 ✓	799.0	801.0	0.060 °F
	1000.0	999.5 ✓	999.9 ✓	999.0	1001.0	0.060 °F
	1200.0	1199.5 ✓	1200.0 ✓	1198.8	1201.2	0.060 °F
	1400.0	1399.5 ✓	1399.8 ✓	1398.6	1401.4	0.060 °F
	1600.0	1599.4 ✓	1599.8 ✓	1598.4	1601.6	0.060 °F
	1800.0	1799.2 ✓	1799.8 ✓	1798.2	1801.8	0.060 °F
	2000.0	1999.2 ✓	1999.8 ✓	1998.0	2002.0	0.060 °F
	2200.0	2199.2 ✓	2199.7 ✓	2197.8	2202.2	0.060 °F
	2800.0	2799.3 ✓	2800.0 ✓	2797.2	2802.8	0.063 °F
	3200.0	3199.5 ✓	3200.0 ✓	3196.8	3203.2	0.063 °F
Type S -58° to 3214°F	-50.0	-50.6 ✓	-50.6 ✓	-52.0	-48.0	0.065 °F
	0.0	-0.8 ✓	-0.4 ✓	-2.0	2.0	0.065 °F
	32.0	31.3 ✓	31.6 ✓	30.0	34.0	0.065 °F
	200.0	199.4 ✓	199.8 ✓	198.8	201.2	0.061 °F
	400.0	399.4 ✓	399.7 ✓	398.8	401.2	0.060 °F
	600.0	599.5 ✓	599.7 ✓	599.0	601.0	0.060 °F
	800.0	799.5 ✓	799.7 ✓	799.0	801.0	0.060 °F
	1000.0	999.7 ✓	999.9 ✓	999.0	1001.0	0.061 °F
	1200.0	1199.7 ✓	1199.8 ✓	1198.8	1201.2	0.061 °F
	1400.0	1399.3 ✓	1399.8 ✓	1398.6	1401.4	0.061 °F
	1600.0	1599.4 ✓	1599.8 ✓	1598.4	1601.6	0.061 °F
	1800.0	1799.1 ✓	1799.8 ✓	1798.2	1801.8	0.061 °F
	2000.0	1999.0 ✓	1999.8 ✓	1998.0	2002.0	0.061 °F
	2200.0	2198.8 ✓	2199.6 ✓	2197.8	2202.2	0.061 °F
	2800.0	2799.1 ✓	2799.9 ✓	2797.2	2802.8	0.064 °F
	3200.0	3198.7 ✓	3199.6 ✓	3196.8	3203.2	0.064 °F
Type N 32° to 2570°F	32.0	31.8 ✓	31.8 ✓	31.0	33.0	0.058 °F
	200.0	199.8 ✓	199.8 ✓	199.0	201.0	0.058 °F
	400.0	399.7 ✓	399.7 ✓	399.0	401.0	0.058 °F
	600.0	599.7 ✓	599.9 ✓	599.0	601.0	0.058 °F
	800.0	799.7 ✓	799.9 ✓	799.0	801.0	0.058 °F
	1000.0	999.7 ✓	999.9 ✓	999.0	1001.0	0.059 °F
	1200.0	1199.7 ✓	1200.0 ✓	1198.8	1201.2	0.059 °F
	1400.0	1399.6 ✓	1400.0 ✓	1398.6	1401.4	0.059 °F



# Certificate of Calibration

1000209034

Page 7 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>Thermocouple Measurement</b>						
	1600.0	1599.6 ✓	1600.0 ✓	1598.4	1601.6	0.059 °F
	1800.0	1799.8 ✓	1800.0 ✓	1798.2	1801.8	0.059 °F
	2000.0	1999.6 ✓	2000.1 ✓	1998.0	2002.0	0.059 °F
	2200.0	2199.2 ✓	2199.7 ✓	2197.8	2202.2	0.060 °F
	2370.0	2369.3 ✓	2369.8 ✓	2367.6	2372.4	0.060 °F
Type B 122° to 3276°F	150.0	144.9 ✓	147.7 ✓	126.3	173.7	0.15 °F
	200.0	194.4 ✓	196.3 ✓	176.3	223.7	0.15 °F
	400.0	397.4 ✓	398.3 ✓	396.2	403.8	0.074 °F
	600.0	598.5 ✓	599.0 ✓	596.2	603.8	0.064 °F
	800.0	798.6 ✓	799.5 ✓	798.0	802.0	0.064 °F
	1000.0	999.0 ✓	999.3 ✓	998.0	1002.0	0.062 °F
	1200.0	1199.3 ✓	1199.8 ✓	1198.0	1202.0	0.062 °F
	1400.0	1399.3 ✓	1399.8 ✓	1398.6	1401.4	0.060 °F
	1600.0	1599.4 ✓	1599.6 ✓	1598.4	1601.6	0.060 °F
	1800.0	1799.2 ✓	1799.6 ✓	1798.2	1801.8	0.060 °F
	2000.0	1999.2 ✓	1999.6 ✓	1998.0	2002.0	0.060 °F
	2200.0	2199.2 ✓	2199.4 ✓	2197.8	2202.2	0.060 °F
	2600.0	2599.5 ✓	2599.7 ✓	2597.4	2602.6	0.060 °F
	3000.0	2999.3 ✓	2999.7 ✓	2997.0	3003.0	0.061 °F
	3272.0	3271.3 ✓	3271.6 ✓	3268.7	3275.3	0.061 °F
<b>Thermocouple Output</b>						
Type E -328° to 1832°F	-328.0	-328.14 ✓	-328.14 ✓	-329.0	-327.0	0.061 °F
	-200.0	-199.93 ✓	-199.93 ✓	-201.0	-199.0	0.058 °F
	32.0	31.91 ✓	32.02 ✓	31.0	33.0	0.058 °F
	600.0	600.13 ✓	599.99 ✓	599.0	601.0	0.058 °F
	1000.0	1000.22 ✓	1000.00 ✓	999.0	1001.0	0.058 °F
	1832.0	1832.29 ✓	1831.81 ✓	1830.2	1833.8	0.059 °F
Type J -328° to 2192°F	-328.0	-328.11 ✓	-327.95 ✓	-329.0	-327.0	0.058 °F
	-200.0	-200.15 ✓	-200.06 ✓	-201.0	-199.0	0.058 °F
	32.0	31.87 ✓	32.00 ✓	31.0	33.0	0.058 °F
	800.0	800.20 ✓	800.02 ✓	799.0	801.0	0.058 °F
	1600.0	1600.21 ✓	1599.87 ✓	1598.4	1601.6	0.059 °F
	2192.0	2192.49 ✓	2191.95 ✓	2189.8	2194.2	0.060 °F
Type K -328° to 2498°F	-328.0	-327.96 ✓	-327.73 ✓	-329.0	-327.0	0.079 °F



# Certificate of Calibration

1000209034

Page 8 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>Thermocouple Output</b>						
	-200.0	-199.98 ✓	-199.86 ✓	-201.0	-199.0	0.058 °F
	32.0	31.86 ✓	32.02 ✓	31.0	33.0	0.058 °F
	600.0	600.13 ✓	600.09 ✓	599.0	601.0	0.058 °F
	1600.0	1600.25 ✓	1599.85 ✓	1598.4	1601.6	0.059 °F
	2498.0	2498.56 ✓	2497.86 ✓	2495.5	2500.5	0.062 °F
Type T -328° to 752°F	-328.0	-327.89 ✓	-327.87 ✓	-329.0	-327.0	0.066 °F
	-200.0	-199.89 ✓	-199.89 ✓	-201.0	-199.0	0.059 °F
	32.0	31.86 ✓	32.01 ✓	31.0	33.0	0.058 °F
	200.0	199.81 ✓	199.90 ✓	199.0	201.0	0.058 °F
	400.0	399.88 ✓	399.89 ✓	399.0	401.0	0.058 °F
	752.0	752.04 ✓	751.87 ✓	751.0	753.0	0.058 °F
Type R -58° to 3214°F	-50.0	-51.29 ✓	-49.65 ✓	-52.6	-47.4	0.065 °F
	32.0	30.92 ✓	32.13 ✓	29.4	34.6	0.065 °F
	600.0	599.56 ✓	600.01 ✓	599.0	601.0	0.060 °F
	1600.0	1599.89 ✓	1599.94 ✓	1598.4	1601.6	0.060 °F
	2200.0	2200.06 ✓	2199.92 ✓	2197.8	2202.2	0.060 °F
	3200.0	3200.31 ✓	3199.68 ✓	3196.8	3203.2	0.063 °F
Type S -58° to 3214°F	-50.0	-51.18 ✓	-49.63 ✓	-52.0	-48.0	0.065 °F
	32.0	30.88 ✓	32.11 ✓	30.0	34.0	0.065 °F
	600.0	599.52 ✓	599.99 ✓	599.0	601.0	0.060 °F
	1600.0	1599.75 ✓	1599.85 ✓	1598.4	1601.6	0.061 °F
	2200.0	2199.92 ✓	2199.78 ✓	2197.8	2202.2	0.061 °F
	3200.0	3200.54 ✓	3199.87 ✓	3196.8	3203.2	0.064 °F
Type N 32° to 2570°F	32.0	31.79 ✓	32.04 ✓	31.0	33.0	0.058 °F
	600.0	600.19 ✓	600.17 ✓	599.0	601.0	0.058 °F
	1400.0	1400.13 ✓	1399.89 ✓	1398.6	1401.4	0.059 °F
	1600.0	1600.27 ✓	1599.93 ✓	1598.4	1601.6	0.059 °F
	2000.0	2000.12 ✓	1999.71 ✓	1998.0	2002.0	0.059 °F
	2370.0	2370.52 ✓	2369.95 ✓	2367.6	2372.4	0.060 °F
Type B 122° to 3276°F	150.0	140.5 ✓	149.1 ✓	126.3	173.7	0.15 °F
	1200.0	1199.05 ✓	1199.79 ✓	1198.0	1202.0	0.062 °F
	1600.0	1599.53 ✓	1599.94 ✓	1598.4	1601.6	0.060 °F
	2000.0	1999.81 ✓	2000.05 ✓	1998.0	2002.0	0.060 °F
	2600.0	2599.93 ✓	2599.97 ✓	2597.4	2602.6	0.060 °F





# Certificate of Calibration

1000209034

Page 9 of 9



✓ In Tolerance    ✗ Out of Tolerance

## Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
<b>Thermocouple Output</b>						
	3272.0	3272.07	✓	3271.78	✓	3268.7    3275.3    0.061 °F

End of Datasheet

## Calibration Standards

<u>NIST Traceable #</u>	<u>Instrument ID#</u>	<u>Description</u>	<u>Model</u>	<u>Calibration Date</u>	<u>Date Due</u>
1000202449	00266	High Impedance Voltmeter-Null Detector	845AR	01 MAY 2018	30 APR 2019
1000202460	00522	Resistance Standard	RS925A	13 JUN 2018	30 APR 2019
1000202912	02037	Humidity/Temperature Chart Recorder	RH520	06 JUN 2018	31 MAY 2019
1000205362	00890	Reference Multimeter	8508A-01	13 NOV 2018	30 NOV 2019
1000207011	00888	Calibrator	5720A	01 DEC 2018	31 DEC 2019