



Certificate of Calibration

1000223607

Page 1 of 9



Customer Information

Accurate Calibration & Repair
1924 Pinnacle Drive
Aurora, IL 60504

PO #: 2021002
Reference #: 2150268Rg17025
Account #: 00317
SO #: 50268

Instrument Identification

Instrument Id: **02.20.002725**

Location:

Noun: Handheld Calibrator

Model: Memocal 2000

Mfr: ERO Electronic

Serial #: 02.20.002725

Accuracy: See manufacturer's specifications

Certification Information

Reason For Service: Calibration with Data
Type Of Calibration: Accredited 17025
As Found Condition: In Tolerance
As Left Condition: Left As Found
Procedure: MFR Manual :

Technician: Appealle Bullock
Cal Date: 27 MAY 21
Cal Due: 27 AUG 21
Temperature: 23.0 °C
Humidity: 44.0 %

Technician Remarks: Datasheet reflects customer specified calibration points. Unit calibrated per AMS 2750 F.

- No sampling plan or other procedure was used for this calibration. Measurements and information on this certificate are valid at time of calibration only and any number of factors may cause calibration to drift out of tolerance prior to calibration due date.
- The calibration results published in this certificate were obtained using equipment capable of producing results that are traceable through NIST to the International System of Units (SI).
- The expanded uncertainty of the measurement process has not exceeded 25% of the tolerance allowed for the individual characteristics measured, unless otherwise stated. The uncertainties are based on a 95% confidence level, K=2.
- (!) Designates that the expanded uncertainty of measurement does not meet the 95% confidence level.
- J.H. Metrology Co., Inc's Calibration Control System complies with applicable requirements of ANSI Z540-1-1994, ISO 9001, and ISO/IEC 17025-2005.
- The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without the written approval of J.H. Metrology Co., Inc.

Approved By:

President

Date: May 27, 2021

Calibration Data

✓ In Tolerance ✗ Out of Tolerance

Range	Nominal	As Found	As Left	Min	Max	Uncertainty
mV Output						
-4 mV to 20 mV	-4.000	-3.99987	✓ As Found	-4.004	-3.996	0.00060 mV
	0.000	0.00034	✓ As Found	-0.004	0.004	0.00059 mV



Certificate of Calibration

1000223607

Page 2 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left		Min	Max	Uncertainty
mV Output								
	5.000	4.99982	✓	As Found		4.996	5.004	0.00060 mV
	10.000	9.99962	✓	As Found		9.996	10.004	0.00061 mV
	15.000	14.99975	✓	As Found		14.996	15.004	0.00062 mV
	20.000	19.99989	✓	As Found		19.996	20.004	0.00063 mV
-40 mV to 200 mV	-40.00	-39.99758	✓	As Found		-40.04	-39.96	0.0058 mV
	0.00	0.00491	✓	As Found		-0.04	0.04	0.0058 mV
	50.00	49.99856	✓	As Found		49.96	50.04	0.0058 mV
	100.00	99.99704	✓	As Found		99.96	100.04	0.0059 mV
	150.00	149.99437	✓	As Found		149.96	150.04	0.0059 mV
	200.00	199.9972	✓	As Found		199.96	200.04	0.0060 mV
-400 mV to 2000 mV	-400.0	-400.0118	✓	As Found		-400.4	-399.4	0.058 mV
	0.0	0.00631	✓	As Found		-0.4	0.4	0.058 mV
	500.0	499.9507	✓	As Found		499.6	500.4	0.058 mV
	1000.0	999.9161	✓	As Found		999.6	1000.4	0.058 mV
	1500.0	1499.8458	✓	As Found		1499.6	1500.4	0.059 mV
	2000.0	1999.8814	✓	As Found		1999.6	2000.4	0.060 mV
-4000 mV to 20000 mV	-4000	-3999.803	✓	As Found		-4005	-3995	0.58 mV
	0	0.41279	✓	As Found		-5	5	0.58 mV
	5000	5000.009	✓	As Found		4995	5005	0.58 mV
	10000	9999.364	✓	As Found		9995	10005	0.58 mV
	15000	14999.419	✓	As Found		14995	15005	0.59 mV
	20000	19999.54	✓	As Found		19995	20005	0.61 mV
mV Measurement								
-20 mV to 20 mV	-20.000	-20.000	✓	As Found		-20.006	-19.994	0.00090 mV
	0.000	-0.001	✓	As Found		-0.006	0.006	0.00077 mV
	5.000	5.000	✓	As Found		4.994	5.006	0.00080 mV
	10.000	9.999	✓	As Found		9.994	10.006	0.00083 mV
	15.000	15.000	✓	As Found		14.994	15.006	0.00086 mV
	20.000	20.000	✓	As Found		19.994	20.006	0.00090 mV
-200 mV to 200 mV	-200.00	-199.98	✓	As Found		-200.06	-199.94	0.0063 mV
	0.00	0.00	✓	As Found		-0.06	0.06	0.0058 mV
	50.00	50.00	✓	As Found		49.94	50.06	0.0059 mV
	100.00	99.99	✓	As Found		99.94	100.06	0.0060 mV
	150.00	149.99	✓	As Found		149.94	150.06	0.0061 mV

✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left		Min	Max	Uncertainty
mV Measurement								
	200.00	200.00	✓	As Found		199.94	200.06	0.0063 mV
-2000 mV to 2000 mV	-2000.0	-1999.8	✓	As Found		-2000.6	-1999.4	0.060 mV
	0.0	0.0	✓	As Found		-0.6	0.6	0.058 mV
	500.0	499.9	✓	As Found		499.4	500.6	0.058 mV
	1000.0	1000.0	✓	As Found		999.4	1000.6	0.059 mV
	1500.0	1500.0	✓	As Found		1499.4	1500.6	0.059 mV
	2000.0	2000.0	✓	As Found		1999.4	2000.6	0.060 mV
-20000 mV to 20000 mV	-20000	-19998	✓	As Found		-20008	-19992	0.59 mV
	0	0	✓	As Found		-8	8	0.58 mV
	5000	5000	✓	As Found		4992	5008	0.58 mV
	10000	9999	✓	As Found		9992	10008	0.58 mV
	15000	14999	✓	As Found		14992	15008	0.59 mV
	20000	19999	✓	As Found		19992	20008	0.59 mV
mA Output								
0 mA to 21 mA	0.000	0.00016985	✓	As Found		-0.003	0.003	0.00058 mA
	1.000	0.9999179	✓	As Found		0.997	1.003	0.00058 mA
	5.000	5.000424	✓	As Found		4.997	5.003	0.00060 mA
	10.000	9.999788	✓	As Found		9.997	10.003	0.00062 mA
	15.000	14.999406	✓	As Found		14.997	15.003	0.00066 mA
	21.000	20.99947	✓	As Found		20.997	21.003	0.00072 mA
mA Measurement								
-20 mA to 20 mA	-20.000	-19.999	✓	As Found		-20.006	-19.994	0.0011 mA
	0.000	0.000	✓	As Found		-0.006	0.006	0.00058 mA
	5.000	4.999	✓	As Found		4.994	5.006	0.00063 mA
	10.000	9.998	✓	As Found		9.994	10.006	0.00074 mA
	15.000	14.999	✓	As Found		14.994	15.006	0.00087 mA
	20.000	19.999	✓	As Found		19.994	20.006	0.0011 mA
-130 mA to 130 mA	-130.00	-130.00	✓	As Found		-130.05	-129.95	0.0094 mA
	0.00	0.00	✓	As Found		-0.05	0.05	0.0058 mA
	10.00	10.01	✓	As Found		9.95	10.05	0.0058 mA
	50.00	50.00	✓	As Found		49.95	50.05	0.0067 mA
	100.00	100.00	✓	As Found		99.95	100.05	0.0082 mA
	130.00	130.00	✓	As Found		129.95	130.05	0.0094 mA



Certificate of Calibration

1000223607

Page 4 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left	Min	Max	Uncertainty
RTD Pt. 100 Ohm Measurement (-200° to 850°C)							
20.68 Ohms	-195.0	-195.0	✓	As Found	-195.6	-194.4	0.058 °C
60.26 Ohms	-100.0	-100.1	✓	As Found	-100.6	-99.4	0.058 °C
100.00 Ohms	0.0	-0.2	✓	As Found	-0.6	0.6	0.058 °C
138.51 Ohms	100.0	99.9	✓	As Found	99.4	100.6	0.058 °C
280.98 Ohms	500.0	499.9	✓	As Found	499.4	500.6	0.059 °C
389.02 Ohms	845.0	844.8	✓	As Found	844.1	845.9	0.060 °C
Ohms Measurement							
0 to 800 Ohms	0.0	0.0	✓	As Found	-0.2	0.2	0.058 Ohms
	10.0	10.0	✓	As Found	9.8	10.2	0.058 Ohms
	100.0	99.9	✓	As Found	99.8	100.2	0.058 Ohms
	300.0	300.0	✓	As Found	299.8	300.2	0.058 Ohms
	500.0	499.9	✓	As Found	499.8	500.2	0.058 Ohms
	790.0	789.9	✓	As Found	789.8	790.2	0.059 Ohms
Thermocouple Measurement							
Type E -328° to 1832°F	-328.0	-328.0	✓	As Found	-329.0	-327.0	0.061 °F
	-200.0	-200.0	✓	As Found	-201.0	-199.0	0.058 °F
	0.0	0.0	✓	As Found	-1.0	1.0	0.058 °F
	32.0	32.0	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.9	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.9	✓	As Found	399.0	401.0	0.058 °F
	600.0	600.1	✓	As Found	599.0	601.0	0.058 °F
	800.0	800.1	✓	As Found	799.0	801.0	0.058 °F
	1000.0	1000.0	✓	As Found	999.0	1001.0	0.058 °F
	1200.0	1200.0	✓	As Found	1198.8	1201.2	0.058 °F
	1400.0	1400.0	✓	As Found	1398.6	1401.4	0.059 °F
	1600.0	1600.0	✓	As Found	1598.4	1601.6	0.059 °F
	1832.0	1832.2	✓	As Found	1830.2	1833.8	0.059 °F
Type J -328° to 2192°F	-328.0	-328.0	✓	As Found	-329.0	-327.0	0.058 °F
	-200.0	-200.0	✓	As Found	-201.0	-199.0	0.058 °F
	0.0	0.0	✓	As Found	-1.0	1.0	0.058 °F
	32.0	32.0	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.8	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.9	✓	As Found	399.0	401.0	0.058 °F
	600.0	600.1	✓	As Found	599.0	601.0	0.058 °F



Certificate of Calibration

1000223607

Page 5 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left	Min	Max	Uncertainty
Thermocouple Measurement							
	800.0	800.1	✓	As Found	799.0	801.0	0.058 °F
	1000.0	999.9	✓	As Found	999.0	1001.0	0.059 °F
	1200.0	1200.0	✓	As Found	1198.8	1201.2	0.059 °F
	1600.0	1600.0	✓	As Found	1598.4	1601.6	0.059 °F
	1800.0	1800.0	✓	As Found	1798.2	1801.8	0.059 °F
	2192.0	2192.2	✓	As Found	2189.8	2194.2	0.060 °F
Type K -328° to 2498°F	-328.0	-328.0	✓	As Found	-329.0	-327.0	0.079 °F
	-200.0	-200.0	✓	As Found	-201.0	-199.0	0.058 °F
	0.0	0.0	✓	As Found	-1.0	1.0	0.058 °F
	32.0	32.0	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.9	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.7	✓	As Found	399.0	401.0	0.058 °F
	600.0	599.9	✓	As Found	599.0	601.0	0.058 °F
	800.0	799.9	✓	As Found	799.0	801.0	0.058 °F
	1000.0	1000.0	✓	As Found	999.0	1001.0	0.059 °F
	1200.0	1200.2	✓	As Found	1198.8	1201.2	0.059 °F
	1400.0	1400.2	✓	As Found	1398.6	1401.4	0.059 °F
	1600.0	1600.0	✓	As Found	1598.4	1601.6	0.059 °F
	1800.0	1800.0	✓	As Found	1798.2	1801.8	0.059 °F
	2000.0	2000.1	✓	As Found	1998.0	2002.0	0.059 °F
	2200.0	2199.9	✓	As Found	2197.8	2202.2	0.062 °F
	2498.0	2498.0	✓	As Found	2495.5	2500.5	0.062 °F
Type T -328° to 752°F	-328.0	-328.0	✓	As Found	-329.0	-327.0	0.066 °F
	-200.0	-200.0	✓	As Found	-201.0	-199.0	0.059 °F
	0.0	0.0	✓	As Found	-1.0	1.0	0.058 °F
	32.0	32.0	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.9	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.9	✓	As Found	399.0	401.0	0.058 °F
	600.0	600.1	✓	As Found	599.0	601.0	0.058 °F
	752.0	752.1	✓	As Found	751.0	753.0	0.058 °F
Type R -58° to 3214°F	-50.0	-50.6	✓	As Found	-52.6	-47.4	0.065 °F
	0.0	-0.8	✓	As Found	-2.6	2.6	0.065 °F
	32.0	31.6	✓	As Found	29.4	34.6	0.065 °F
	200.0	199.8	✓	As Found	198.4	201.6	0.061 °F
	400.0	399.7	✓	As Found	398.8	401.2	0.060 °F



Certificate of Calibration

1000223607

Page 6 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left	Min	Max	Uncertainty
Thermocouple Measurement							
	600.0	599.9	✓	As Found	599.0	601.0	0.060 °F
	800.0	799.9	✓	As Found	799.0	801.0	0.060 °F
	1000.0	999.9	✓	As Found	999.0	1001.0	0.060 °F
	1200.0	1200.0	✓	As Found	1198.8	1201.2	0.060 °F
	1400.0	1399.9	✓	As Found	1398.6	1401.4	0.060 °F
	1600.0	1599.8	✓	As Found	1598.4	1601.6	0.060 °F
	1800.0	1799.8	✓	As Found	1798.2	1801.8	0.060 °F
	2000.0	1999.8	✓	As Found	1998.0	2002.0	0.060 °F
	2200.0	2199.9	✓	As Found	2197.8	2202.2	0.060 °F
	2800.0	2800.0	✓	As Found	2797.2	2802.8	0.063 °F
	3200.0	3200.0	✓	As Found	3196.8	3203.2	0.063 °F
Type S -58° to 3214°F	-50.0	-50.6	✓	As Found	-52.0	-48.0	0.065 °F
	0.0	0.0	✓	As Found	-2.0	2.0	0.065 °F
	32.0	31.8	✓	As Found	30.0	34.0	0.065 °F
	200.0	199.8	✓	As Found	198.8	201.2	0.061 °F
	400.0	399.7	✓	As Found	398.8	401.2	0.060 °F
	600.0	599.9	✓	As Found	599.0	601.0	0.060 °F
	800.0	799.7	✓	As Found	799.0	801.0	0.060 °F
	1000.0	999.9	✓	As Found	999.0	1001.0	0.061 °F
	1200.0	1200.0	✓	As Found	1198.8	1201.2	0.061 °F
	1400.0	1399.8	✓	As Found	1398.6	1401.4	0.061 °F
	1600.0	1599.8	✓	As Found	1598.4	1601.6	0.061 °F
	1800.0	1799.8	✓	As Found	1798.2	1801.8	0.061 °F
	2000.0	1999.8	✓	As Found	1998.0	2002.0	0.061 °F
	2200.0	2199.6	✓	As Found	2197.8	2202.2	0.061 °F
	2800.0	2800.0	✓	As Found	2797.2	2802.8	0.064 °F
	3200.0	3199.8	✓	As Found	3196.8	3203.2	0.064 °F
Type N 32° to 2570°F	32.0	31.8	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.8	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.7	✓	As Found	399.0	401.0	0.058 °F
	600.0	599.9	✓	As Found	599.0	601.0	0.058 °F
	800.0	799.9	✓	As Found	799.0	801.0	0.058 °F
	1000.0	999.9	✓	As Found	999.0	1001.0	0.059 °F
	1200.0	1200.0	✓	As Found	1198.8	1201.2	0.059 °F
	1400.0	1400.0	✓	As Found	1398.6	1401.4	0.059 °F



Certificate of Calibration

1000223607

Page 7 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left	Min	Max	Uncertainty
Thermocouple Measurement							
	1600.0	1600.0	✓	As Found	1598.4	1601.6	0.059 °F
	1800.0	1800.1	✓	As Found	1798.2	1801.8	0.059 °F
	2000.0	2000.1	✓	As Found	1998.0	2002.0	0.059 °F
	2200.0	2200.1	✓	As Found	2197.8	2202.2	0.060 °F
	2370.0	2370.0	✓	As Found	2367.6	2372.4	0.060 °F
Type B 122° to 3276°F	150.0	150.6	✓	As Found	126.3	173.7	0.15 °F
	200.0	198.1	✓	As Found	176.3	223.7	0.15 °F
	400.0	399.0	✓	As Found	396.2	403.8	0.074 °F
	600.0	599.5	✓	As Found	596.2	603.8	0.064 °F
	800.0	799.5	✓	As Found	798.0	802.0	0.064 °F
	1000.0	999.3	✓	As Found	998.0	1002.0	0.062 °F
	1200.0	1199.8	✓	As Found	1198.0	1202.0	0.062 °F
	1400.0	1399.8	✓	As Found	1398.6	1401.4	0.060 °F
	1600.0	1599.6	✓	As Found	1598.4	1601.6	0.060 °F
	1800.0	1799.8	✓	As Found	1798.2	1801.8	0.060 °F
	2000.0	1999.6	✓	As Found	1998.0	2002.0	0.060 °F
	2200.0	2199.7	✓	As Found	2197.8	2202.2	0.060 °F
	2600.0	2599.9	✓	As Found	2597.4	2602.6	0.060 °F
	3000.0	2999.8	✓	As Found	2997.0	3003.0	0.061 °F
	3272.0	3271.8	✓	As Found	3268.7	3275.3	0.061 °F
Thermocouple Output							
Type E -328° to 1832°F	-328.0	-328.11	✓	As Found	-329.0	-327.0	0.061 °F
	-200.0	-199.91	✓	As Found	-201.0	-199.0	0.058 °F
	32.0	32.02	✓	As Found	31.0	33.0	0.058 °F
	600.0	599.97	✓	As Found	599.0	601.0	0.058 °F
	1000.0	999.97	✓	As Found	999.0	1001.0	0.058 °F
	1832.0	1831.75	✓	As Found	1830.2	1833.8	0.059 °F
Type J -328° to 2192°F	-328.0	-327.93	✓	As Found	-329.0	-327.0	0.058 °F
	-200.0	-200.06	✓	As Found	-201.0	-199.0	0.058 °F
	32.0	32.00	✓	As Found	31.0	33.0	0.058 °F
	800.0	799.99	✓	As Found	799.0	801.0	0.058 °F
	1600.0	1599.82	✓	As Found	1598.4	1601.6	0.059 °F
	2192.0	2191.86	✓	As Found	2189.8	2194.2	0.060 °F
Type K -328° to 2498°F	-328.0	-327.71	✓	As Found	-329.0	-327.0	0.079 °F



Certificate of Calibration

1000223607

Page 8 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found		As Left	Min	Max	Uncertainty
Thermocouple Output							
	-200.0	-199.85	✓	As Found	-201.0	-199.0	0.058 °F
	32.0	32.00	✓	As Found	31.0	33.0	0.058 °F
	600.0	600.06	✓	As Found	599.0	601.0	0.058 °F
	1600.0	1599.82	✓	As Found	1598.4	1601.6	0.059 °F
	2498.0	2497.77	✓	As Found	2495.5	2500.5	0.062 °F
Type T -328° to 752°F	-328.0	-327.87	✓	As Found	-329.0	-327.0	0.066 °F
	-200.0	-199.89	✓	As Found	-201.0	-199.0	0.059 °F
	32.0	32.00	✓	As Found	31.0	33.0	0.058 °F
	200.0	199.89	✓	As Found	199.0	201.0	0.058 °F
	400.0	399.87	✓	As Found	399.0	401.0	0.058 °F
	752.0	751.85	✓	As Found	751.0	753.0	0.058 °F
Type R -58° to 3214°F	-50.0	-49.74	✓	As Found	-52.6	-47.4	0.065 °F
	32.0	32.05	✓	As Found	29.4	34.6	0.065 °F
	600.0	599.97	✓	As Found	599.0	601.0	0.060 °F
	1600.0	1599.85	✓	As Found	1598.4	1601.6	0.060 °F
	2200.0	2199.79	✓	As Found	2197.8	2202.2	0.060 °F
	3200.0	3199.57	✓	As Found	3196.8	3203.2	0.063 °F
Type S -58° to 3214°F	-50.0	-49.68	✓	As Found	-52.0	-48.0	0.065 °F
	32.0	32.04	✓	As Found	30.0	34.0	0.065 °F
	600.0	599.94	✓	As Found	599.0	601.0	0.060 °F
	1600.0	1599.73	✓	As Found	1598.4	1601.6	0.061 °F
	2200.0	2199.70	✓	As Found	2197.8	2202.2	0.061 °F
	3200.0	3199.68	✓	As Found	3196.8	3203.2	0.064 °F
Type N 32° to 2570°F	32.0	32.02	✓	As Found	31.0	33.0	0.058 °F
	600.0	600.13	✓	As Found	599.0	601.0	0.058 °F
	1400.0	1399.81	✓	As Found	1398.6	1401.4	0.059 °F
	1600.0	1599.87	✓	As Found	1598.4	1601.6	0.059 °F
	2000.0	1999.65	✓	As Found	1998.0	2002.0	0.059 °F
	2370.0	2369.87	✓	As Found	2367.6	2372.4	0.060 °F
Type B 122° to 3276°F	150.0	149.8	✓	As Found	126.3	173.7	0.15 °F
	1200.0	1199.66	✓	As Found	1198.0	1202.0	0.062 °F
	1600.0	1599.89	✓	As Found	1598.4	1601.6	0.060 °F
	2000.0	1999.94	✓	As Found	1998.0	2002.0	0.060 °F
	2600.0	2599.86	✓	As Found	2597.4	2602.6	0.060 °F



Certificate of Calibration

1000223607

Page 9 of 9



✓ In Tolerance ✗ Out of Tolerance

Calibration Data

Range	Nominal	As Found	As Left	Min	Max	Uncertainty	
Thermocouple Output							
	3272.0	3271.71	✓	As Found	3268.7	3275.3	0.061 °F

End of Datasheet

Calibration Standards

<u>NIST Traceable #</u>	<u>Instrument ID#</u>	<u>Description</u>	<u>Model</u>	<u>Calibration Date</u>	<u>Date Due</u>
1000216271	01241	Precision Process Calibrator	7526A	18 JUL 2020	31 JUL 2021
1000220682	00888	Calibrator	5720A	08 JAN 2021	31 JAN 2022
1000220683	00890	Multimeter, 8.5 Digit Reference	8508A-01	26 JAN 2021	31 JAN 2022
1000222775	00266	High Impedance Voltmeter-Null Detector	845AR	22 APR 2021	30 APR 2022
1000222787	00522	Resistance Standard	RS925A	22 APR 2021	30 APR 2022
1000222817	01196	RH/Temperature Data Logger	EL-USB-2-LCD	29 APR 2021	30 APR 2022