



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Accurate Calibration & Repair Service, Inc.
1924 Pinnacle Drive, Aurora, IL 60502

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

System Accuracy Testing of Thermocouples and Temperature Uniformity Survey
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

January 17, 2003

Issue Date:

June 29, 2021

Expiration Date:

September 30, 2023

Accreditation No:

59048

Certificate No:

L21-421

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Accurate Calibration & Repair Service, Inc.

1924 Pinnacle Drive, Aurora, IL 60502

Contact Name: Roy Halters Phone: 630-236-0439

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Thermodynamics ^o	Ovens and Furnaces Freezers and Dry Boxes Thermocouple	Uniformity Survey Thermocouple Type K	AMS 2750 F	200 °F to 2 400 °F D.L. = 0.1 °F/ 0.1 °C
		Uniformity Survey Thermocouple Type T		-320 °F to 200 °F D.L. = 0.1 °F/ 0.1 °C
		System Accuracy		100 °F to 1 900 °F D.L. = 0.1 °F/ 0.1 °C

- The presence of a superscript ^o means that the laboratory performs testing of the indicated parameter onsite at customer locations. Example: Outside Micrometer ^o would mean that the laboratory performs this testing onsite at the customer's location.

